

## Indicator Strips



Indicator strips are to be placed in each pack of each sterilizer load: Strips changes the colour on correct integration of temperature, time and humidity and concentration of sterilizing agent.

- The AC 250/500 steam strip (type 4) changes colour from white to black
- The AC 6 steam strip (type 6) changes colour from green to brown
- The EO strip (type 5) changes colour from purple to green
- The PL - plasma strip (type 1) changes colour from blue to pink

in accordance with ISO 11140-1.

Article	Width	Length	Gusset	Packing Unit
AC250/250	15	200		250
AC6	15	100		250
EO250	15	200		250

Article	Width	Length	Gusset	Packing Unit
PL500	20	150		500